

BE Factor V Deficient Plasma FV

Immuno-depleted plasma for the determination of Factor V activity in human plasma

REF 771605: DP (6 x 1 mL)

INTENDED USE

This reagent is designated for professional use in laboratory (semi-automated or automated method). It allows the quantitative determination of Factor V activity in citrated human plasma to assess the status of coagulation factors normally found in blood.

This test is realized with Behnk reagents as follows:

- REF 771100, REF 771101: BE PT LI Thromboplastin low ISI
- REF 771150, REF 771151: BE PT HI Thromboplastin high ISI
- REF 771700: BE Owren Buffer (Plasma dilution buffer)

PRINCIPLE ⁽¹⁾

The test is based on the measurement of clotting time in the presence of thromboplastin and calcium with a method in which all factors are present in excess (supplied by Factor V Deficient Plasma) except Factor V, which is derived from the sample to be tested.

GENERALITIES ^{(2) (7) (8) (9)}

The structure of platelet factor V is very similar to that of plasma factor V. It is activated by Thrombin.


The Factor Va forms an equimolar complex with the Factor Xa in the presence of Ca²⁺ and phospholipids. This complex activates the Prothrombin into Thrombin.

The Factor Va is inactivated by Factor Xa, activated protein C (PCa) and plasmin. Some mutations of FV (i.e. Leiden) induce a resistance of FVa to its inactivation by PCa which is a risk factor for the development of thrombosis, especially when other risk factors are present.

Pathological variations may be observed in following cases:

- Isolated deficiency of Factor V:
 - Congenital deficiency
 - Acquired deficiency associated to Factor V inhibitors
- Acquired Factor V deficiency associated with deficiencies of other coagulation Factors:
 - Liver diseases: cirrhosis, hepatitis
 - Disseminated intravascular coagulation (DIC)

REAGENTS

DP	FV	Deficient Plasma FV	
Freeze dried citrated plasma without Factor V, removed by selective immune adsorption.			Human Origin

According to 1272/2008 regulation, this reagent is not classified as dangerous.

SAFETY CAUTIONS ^{(10) (11)}

- Refer to current Material Safety Data Sheet available on request or on www.behnk.de
 - Each donor unit used to manufacture this product was tested and found non-reactive for HbsAg, antibody to Hepatitis C and antibody to HIV-1/HIV-2.
 - However, no test method can offer complete assurance that infectious agents are absent. All specimens or reagents from biological origin should be handled as potentially infectious, in accordance with good laboratory practices using appropriate precautions.
 - Waste disposal: Respect legislation in force in the country.
- Any serious incident that has occurred in connection with the device is notified to the manufacturer and the competent authority of the Member State in which the user and/or patient is based.

PREPARATION OF REAGENTS

Open the vial carefully and add exactly the volume of demineralised water stated on the label.

Cap the vial and let stand for 15 minutes at room temperature.

Mix gently by swirling and inverting before use, to homogenise the content.

STABILITY AND STORAGE

Unopened vials, stored away from light at 2-8 °C are stable until the expiry date stated on the label.

Stability after reconstitution:

- 2-8 °C 8 hours
- On board Stability (OBS)* 4 hours
- 15-25 °C 4 hours

* 18-22 °C

Do not use any reagent after expiry date.

SAMPLES COLLECTION AND HANDLING ^{(5) (9)}

Plasma from careful venipuncture with anticoagulant ratio of 1/10 (sodium citrate solution 0.109 M). Mix immediately the blood with anticoagulant.

Avoid drawing with a syringe that could result in the formation of micro-clots.

Centrifuge for 10 minutes at 3000 g and extract supernatant.

Stability:

- 4 h at 20-25 °C, 8 h at 2-8 °C
- 15 days at -20 °C, 1 month at -80 °C (if quickly frozen. Defrost at 37 °C until complete defrosting).

Caution: if the same plasma is used for testing Factor VII, do not store at 2-8°C, because the Factor VII may be activated by the kallikrein system in this temperature range.

LIMITS ⁽³⁾

Thrombin inhibitors (hirudin, argatroban, ...) present in the specimen may decrease Factor V activity in the specimen.

For a more comprehensive review of factors affecting this assay, refer to the publication of Young D.S.

MATERIAL REQUIRED BUT NOT PROVIDED

Basic medical analysis laboratory equipment
Automated or semi-automated coagulation analyzer
Demineralised water

REFERENCE RANGE ⁽⁷⁾

Plasma (adult): Usually > 70 %

Each laboratory should establish its own normal ranges for the population that it serves.

QUALITY CONTROL

REF 773100: BE Trol 1; REF 773101: BE Trol 2

Controls are required for checking the accuracy and reproducibility of the results.

The control intervals should be adapted to each laboratory's individual requirements.

Values obtained should fall within the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

PROCEDURE

Manual method on semi-automated systems

Pre-incubate PT reagent (Thromboplastin) 15 min to reach a temperature of 37 °C and mix gently before use.

Dilute samples and controls: 1/10 in BE Owren Buffer.

Calibrators: prepare dilutions as indicated in § Calibration.

- Diluted Plasma (calibrators, controls, plasmas): 100 µL
- Deficient Plasma: 100 µL

Incubate for 120 sec at 37 °C

- Thromboplastin (37 °C): 200 µL

The automatic countdown timer will start immediately after Thromboplastin addition and stop when the clot is formed.

Automated method on Behnk Thrombolyzer series

Refer to the full detailed application specific to the automated system.

Note:

- Performances and stability data have been validated on Thrombolyzer Compact X (available on request).
- With manual procedure and on other automated coagulation analyzer, performances and stability data must be validated by user.
- Other validated applications or proposal applications are available on request.

CALIBRATION

Use REF 775100: BE Cal Ref
Reference plasma traceable to WHO SSC/ISTH Secondary Coagulation Standard NIBSC code: SSCLOT4.

Manual method on semi-automated analyzer: Prepare a calibration curve with dilution 1/10, 1/20, 1/40 and 1/80 in BE Owren Buffer. Measure in triplicate the clotting time of each level.

Automated method on Behnk Thrombolyzer series: Perform a calibration with BE Cal Ref using automatic dilutions indicated in the specific application.

CALCULATION

Results are expressed in % of Deficient Factor according to the calibration curve.

PERFORMANCES

The studies were performed on Thrombolyzer Compact X.

Precision:

Within run N = 20	Level 1	Level 2	Between run N = 20	Level 1	Level 2
Mean (%)	96	33	Mean (%)	81	47
S.D. (%)	3.0	1.0	S.D. (%)	6.1	2.4
C.V. %	3.6	3.3	C.V. %	7.5	5.2

Detection limit: equivalent to 3 % of Factor V

Measuring Range: from 10 % (QL) to 100 %

Interferences (PT LI, sec):

Turbidity	No interference up to 450 mg/dL of Triglycerides
Low Molecular weight heparin	No interference up to 0.114 IU anti Xa
Unfractionated heparin	No interference up to 0.038 IU anti Xa
Bilirubin	Negative interference from 228 µmol/L
Hemoglobin	No interference up to 258 µmol/L

Other substances may interfere with the results (see § Limits)

Calibration Stability: Make a new calibration when changing reagent batch, if quality control results are found out of the established range and after maintenance operations.

REFERENCES

- (1) SOULIER J.P., LARRIEU M.-J.: *Sang.* 23, 7, 549-559, 1952
- (2) CAEN J., LARRIEU M.-J., SAMAMA M.: *Paris, L'Expansion scientifique*, 153, 347, 1975
- (3) YOUNG D.S., *Effect of Drugs on Clinical laboratory Tests*, 4th Ed. (1995) p.3-254 à 3-257
- (4) SAMAMA M., CONARD J., HORELLOU M.H., LECOMPTE T.: "Physiologie et exploration de l'hémostasie". *PARIS: DOIN*, 81-82, 116-118, 148-149, 1990
- (5) GIOANNES H., FAGERHOL M.K.: *Acta Obstet. Gynecol. Scand.*, 54, 363-367, 1975
- (6) ALEXANDRE P.: *Paris: Edition Marketing Elipses*, 459-464, 1994
- (7) SAMPOL J., ARNOUX D., BOUTIERE B.: "Manuel Hémostasie", *Paris: Editions scientifiques et médicales Elsevier*, 47, 366-368, 383-384, 409-430, 1995
- (8) ZOLLER B. HILLARP A., DAHLBACK B.: "Activated Protein C resistance": *clinical implication" Clin. Appl. Thromb., Hemost.*, 3, 1, 25-32, 1997
- (9) WOODHAMS B., M.C. DUGA S., ASSELT R.n, MALCOVATI M., PEYVANDI F., SANTAGODISTINO E., MANNUCCI P.M., TENCHINI M.L.: *Blood*, 102, 9, 3210-3216, 2003
- (10) *Occupational Safety and Health Standards; Bloodborne pathogens (29CFR1910.1030) Federal Register July 1, (1998) ; 6, p.267-280*
- (11) *Directive du conseil de l'Europe (90/679/CEE) J. O. de la communauté européenne n°L374 du 31.12.1990, p.1-12*

| = Significant modifications

Manufacturer	Expiry date	In Vitro Diagnostic	Temperature limitation	Catalogue number	See insert	Batch number	Store away from light	Sufficient for	Dilute with	Demineralized water	Biological hazard
--------------	-------------	---------------------	------------------------	------------------	------------	--------------	-----------------------	----------------	-------------	---------------------	-------------------